

Nikon SDC - Shut-down and clean up

When you are done imaging, remove your sample from the microscope and place it on a kimwipe or other paper towel if it has oil or water.

If you used an oil objective, please refer to the [instructions](#) of Dr. Dragavon to clean it:

Shut down the system according to the shut-down procedure taped on the wall:

1. Save your beautiful images to your Network drive.
2. Exit Nikon Elements and disconnect from your Network drive
3. Remove your sample from the stage and clean up the microscope

Clean oil from the objective and remove your waste, wipe everything with ethanol

4. Open the reservation calendar

If there is a user within 1 hour of your session, turn off the environmental chamber (if used) and leave the system on.

If the next user is more than 1 hour after your session, turn everything off (step 5).

5. Using the remote control ONLY:

Turn off the Microscope

Turn off the lasers

Turn off the environmental chamber (follow the instructions on the wall)

Revision #7

Created 4 October 2023 22:06:15 by Evolene Premillieu

Updated 22 May 2024 15:43:16 by Evolene Premillieu